

Pomůcka pro identifikaci výrobku

ENDOPATH® Electrosurgery Probe Plus II

Díky této pomůcce mohou zákazníci identifikovat výrobky s uvedeným datem expirace na značení na obalu výrobku. Tento dokument se týká značení na sekundárním obalu (krabice jednotlivého balení), primárním obalu Tyvek® (víko) a terciárním obalu (prodejní balení) u výrobků s těmito kódy:


Úplný název zdravotnického prostředku	Kódy výrobků
EPS01-EPS08 s datem expirace od září 2015 do března 2018	
ENDOPATH® Electrosurgery Probe Plus II Monopolar Electrosurgery Suction and Irrigation Shaft with Hook Electrode – 5 mm	EPS01
ENDOPATH® Electrosurgery Probe Plus II Monopolar Electrosurgery Suction and Irrigation Shaft with Spatula Electrode – 5 mm	EPS02
ENDOPATH® Electrosurgery Probe Plus II Monopolar Electrosurgery Suction and Irrigation Shaft with Right Angle Electrode – 5 mm	EPS03
ENDOPATH® Electrosurgery Probe Plus II Monopolar Electrosurgery Suction and Irrigation Shaft with Curved Dissector Electrode – 5 mm	EPS04
ENDOPATH® Electrosurgery Probe Plus II Monopolar Electrosurgery Suction and Irrigation Shaft with Hook Electrode – 5 mm	EPS05
ENDOPATH® Electrosurgery Probe Plus II Monopolar Electrosurgery Suction and Irrigation Shaft with Spatula Electrode – 5 mm	EPS06
ENDOPATH® Electrosurgery Probe Plus II Monopolar Electrosurgery Suction and Irrigation Shaft with Right Angle Electrode – 5 mm	EPS07
ENDOPATH® Electrosurgery Probe Plus II Monopolar Electrosurgery Suction and Irrigation Shaft with Curved Dissector Electrode – 5 mm	EPS08
EPS10-EPS13 s datem expirace od dubna April 2013 do března 2018	
ENDOPATH® Electrosurgery Probe Plus II Pool/Sump Suction and Irrigation Shaft – 10 mm	EPS10
ENDOPATH® Electrosurgery Probe Plus II Suction and Irrigation Shaft – 10 mm	EPS11
ENDOPATH® Electrosurgery Probe Plus II Stone Retrieval Shaft – 10 mm	EPS12
ENDOPATH® Electrosurgery Probe Plus II Suction and Irrigation Shaft with Accessory Port – 10 mm	EPS13

Výrobek EPS01 slouží jako příklad.

KRABÍČKA PRODEJNÍHO BALENÍ


(sekundární značení)

ČELNÍ STĚNA a STĚNA DÁVKOVAČE



ZNAČENÍ PRODEJNÍHO BALENÍ

(nachází se na stěně jednotlivého kartonu)



Datum expirace

**Dotčený výrobek:
Viz výše uvedená
tabulka**

ZNAČENÍ NA PŘEPRAVNÍM OBALU




Kód výrobku

Datum expirace

Dotčený výrobek:
Viz výše uvedená
tabulka

Jednokusové balení Tyvek®



Datum
expirace

Dotčený
výrobek:
Viz výše
uvedená
tabulka

Kód
výrobku